

**CERTIFICATE**
FOR EUROPEAN PRODUCT SAFETY

Reference No. SE-S-2300984

High Current Multiplexer

Type designation: RM 16061

Certificate holder: National Instruments Corporation
11500 N Exy,
Austin, TX 78759, United States of America

The product complies with the standard(s): EN 61010-1:2010+A1

Date of expiry: 26 September, 2028

EU Directive information: According to the principle of presumption of conformity, this certificate constitutes support for an EU Declaration of Conformity to the Low Voltage Directive 2014/35/EU.

Additional information in Appendix

Certification Body: Intertek Semko AB, Product Certification **Place:** Kista - Stockholm

Signed:  **Date:** 26 September, 2023

Fredrik Wennersten

APPENDIX:

Reference No. SE-S-2300984

Technical Data:

Type designation RM 16061
Input Ratings: Power: 24VDC, 3A
Charge: 20VDC, 10A
Bias: 230V, 3A

Output Ratings: Max Pulse Current: 1200A
Max Pulse duration: 7ms
Max Bias Voltage: 230V
Max Bias Current: 3A
Pulse/Load: 20 VDC, 1200A

Trademark



Test Report 105382303DAL-002
Product information -

Manufacturing Sites: NI Hungary KFT
Hatar ut 1/A
4031 Debrecen Hungary

According to the principle of presumption of conformity, this certificate, which includes production control, constitutes support for an EU Declaration of Conformity to the Low Voltage Directive 2014/35/EU. This presumption can expire before the end of validity of this certificate due to new issued Standard or Amendment and changes within the EU legislation.

" On-site verification of the production line testing according to Annex F was excluded."

Certification Body: Intertek Semko AB, Product Certification **Place:** Kista - Stockholm

Signed:  **Date:** 26 September, 2023

Fredrik Wennersten