


Intertek **CERTIFICATE**
FOR EUROPEAN PRODUCT SAFETY

Reference No. SE-S-2500095

Field Diagnostic Tester

Type designation: FDT-16050

Certificate holder: National Instruments Corporation
11500 N. Mopac Expy
78759 Austin, TX United States of America


The product complies with the standard(s): EN 61010-1:2010+A1,
EN IEC 61010-2-030:2021 + A11

Date of expiry: 07 February, 2030

EU Directive information: The product listed has been evaluated by a Third-Party Body and deemed to meet the listed harmonized standard(s). This Certificate can be used in support of a claim in meeting relevant EU legislation and mandatory Conformity Marking. And in accordance with EU law, the claim is the sole obligation of the Manufacturer/ Importer.

Additional information in Appendix

Certification Body: Intertek Semko AB, Product Certification **Place:** Kista - Stockholm

Signed: 
Anneli Averland Johansson **Date:** 07 February, 2025
AMP

APPENDIX:

Reference No. SE-S-2500095

Technical Data:

Type designation
Rating

FDT-16050
J2 Connector: 24Vdc, 48W Max.
J9 Connector: 1500Vdc (differential), 800Vdc Ch-Earth CAT I 0Vpk
Transient; 1.5W
J11/J12 Connector: 250Vdc (differential), 250Vdc Ch-Earth CAT I 0Vpk
Transient; 0.25W
All Other Connectors: 60Vdc Max., 5W Max.

Trademark



or "NI"

Test Report

105852622DAL-001,
105852622DAL-002

Product information

-

Manufacturing Sites:

NI Hungary Kft.
4031 Debrecen, Hatar ut 1/A,
Hungary


" On-site verification of the production line testing according to Annex F was excluded."

Certification Body:

Intertek Semko AB, Product Certification

Place: Kista - Stockholm

Signed:



Date: 07 February, 2025

Anneli Averland Johansson

Intertek